



# CMA Testing and Certification Laboratories

廠商會檢定中心

## TEST REPORT

Report No : AZ0014565(0) Date: 14 Apr 2020

Application No : LZ007627(0)

Applicant : SHAN HO ASIA PACIFIC LIMITED  
1/F, 3 DAI KWAI STREET  
TAI PO INDUSTRIAL ESTATE  
NEW TERRITORIES, HONG KONG

Sample Description : One (1) submitted sample stated to be Surgical Mask.  
Sample Status Upon Receipt : Room Temperature



Date Received : 07 Apr 2020

Test Period : 07 Apr 2020 to 09 Apr 2020

Test Requested : Total Bacterial Count 細菌菌落總數

Test Method : With reference to GB 15979-2002, Appendix B.

Test Result : Refer to the result(s) on page 2.

For and on behalf of  
CMA Industrial Development Foundation Limited

Authorized Signature : \_\_\_\_\_

Choy Oi Lan, Ellen  
Assistant Manager

Page 1 of 2



# CMA Testing and Certification Laboratories

廠商會檢定中心

## TEST REPORT

Report No : AZ0014565(0)

Date: 14 Apr 2020

Application No : LZ007627(0)

Test Result :

Test Item	Result	Specification
Total Bacterial Count 細菌菌落總數(CFU/g)	<20	≤ 200

- Note :
1. CFU/g denotes colony forming unit per gram.
  2. The specification is referred to *GB15979-2002 Hygienic Standard for Disposable Sanitary Products*: 口罩 (普通級).

\*\*\*\*\* End of Report \*\*\*\*\*